

Perle IOLAN MDC Medical Terminal Servers

CE Declaration of Conformity

Manufacture's Name: **Perle Systems Limited**

Manufacture's Address: 60 Renfrew Drive
Markham, Ontario
L3R 0E1
Canada

The Directives covered by this Declaration:

2004/108/EC Electromagnetic Compatibility Directive
93/42/EEC (MDD) Medical Device Directive

The Products Covered by this Declaration:

IOLAN MDC Medical Terminal Servers

The Basis on which Conformity is being declared:

The product identified above complies with the requirements of the EMC Directive above by meeting the following standard:

EN 55022:1998 + A1:2000 + A2:2003 Class A
EN 61000-3-2:2000
EN 61000-3-3:1995
EN 55024:1998+A1:2001 +A2:2003

The product identified above complies with the requirements of the Medical Device Directive above by meeting the following standards:

EN 60601-2:2001, Immunity for Non Life-Supporting Equipment
IEC 60601-1:1988+A1:1991+A2:1995 (safety)

The technical documentation required to demonstrate that the product meets the requirements of the Medical Device Directive & Electromagnetic Compatibility Directive has been compiled by the signatory below and is available for inspection by the relevant enforcement authorities.

The products described above comply with the essential requirements of the directives specified.

Approved by:
John Feeney, Chief Operating Officer

Signed

Date: 21st April 2010

