

**Perle IOLAN Medical Terminal Servers  
CE Declaration of Conformity**

Manufacture's Name: **Perle Systems Limited**  
Manufacture's Address: **60 Renfrew Drive  
Markham, Ontario  
Canada L3R 0E1**

The Directives covered by this Declaration:  
**2004/108/EC Electromagnetic Compatibility Directive**  
**93/42/EEC (MDD) Medical Device Directive**  
**2011/65/EU RoHS2 Directive**

The Products Covered by this Declaration:

**IOLAN Medical Terminal Servers**

The basis on which Conformity is being declared:  
The product identified above complies with the requirements of the EMC Directive above by meeting the following standard:

**CISPR 22:2008/EN55022:2010+AC2011 (Class A)**  
**CISPR 24:2010/EN 55024:2010**

The product identified above complies with the requirements of the Medical Device Directive above by meeting the following standards:

**EN 60601-2:2001, Immunity for Non Life-Supporting Equipment**  
**IEC 60601-1:1988+A1:1991+A2:1995 (safety)**

The products identified above are in conformity with Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment and meet the following standard:

**EN50581:2012**

The technical documentation required to demonstrate that the product meets the requirements of the Medical Device Directive, Electromagnetic Compatibility Directive and RoHS2 Directive has been compiled by the signatory below and is available for inspection by the relevant enforcement authorities.

The products described above comply with the essential requirements of the directives specified.

Approved by:  
John Feeney, Chief Operating Officer

Signed



Date: 11<sup>th</sup> February, 2015